

America

CERTIFICATE

No. U8 042849 0011 Rev. 00

Holder of Certificate: **Leica Microsystems AG**
9435 Heerbrugg
SWITZERLAND

Certification Mark:



C

US

Product: **Laboratory Equipment
(Microscopes)**


This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713263234

Date, 2023-10-19

Siemon

(Thorsten Siemon)



America

CERTIFICATE

No. U8 042849 0011 Rev. 00

Model(s): EZ4
EZ4 E
EZ4 W

Brand Name: Leica

Tested according to: UL 61010-1:2012/R:2019-07
CSA C22.2 No. 61010-1:2012/A1:2018-11

Parameters:

Voltage: 100-240 VAC
Frequency: 50 / 60 Hz
Wattage: max. 15 W for EZ4,
max. 30 W for EZ4 E, EZ4 W

Remarks/CoA:

- The EUT shall comply with the National Standards and/or Electrical Codes of the country in question.
- This equipment is for indoor-use in non-hazardous locations.
- The detachable power cord set shall comply with the National Standards and/or National Regulations of the country in question and/or in compliance with this standard.
- The detachable power cord set length shall be not more than 3 m.